

PARIPALLY, KOLLAM-691574

Telephone Office: 0474 2572574, Fax: 0474 2572579 Telephone Principal: 0474-2572572 Electronic mail: gmckollam@gmail.com

Section Email: acts.gmckollam@gmail.com

No. C1/9487/2023/GMC-KLM

Dated: 12/12/2023

QUOTATION NOTICE

Quotation Number	C2/9487/2023/GMC-KLM Dated:15/11/2023		
Due date and time for receipt of quotations	At 2.30 pm on- 29 /12/2023		
Date and time for opening of quotations	At 3.00 pm on 29 /12/2023		
Date up to which the rates are to remain firm for acceptance	31-03-2024		
Designation and address of officer to whom the quotation is to be addressed	Principal Govt. Medical College, Kollam -691 574		

Purchase of material sanction under Annual Plan 2023-2024under M&E(NON KMSCL)

(ITEMS ARE IN ANNEXURE)

* Rate should be quoted in the following format

SI. No	Name Item	Unit cost	% of GST & Rate	Unit cost including GST	Qty.	Total cost Including GST

Sealed quotations are invited for the supply of the materials specified in the schedule attached below/overleaf. The rates quoted should be for delivery of the articles at the place mentioned below the schedule. The necessary superscription, due date for the receipt of quotations, the date up to which the rates will have to remain firm for acceptance and the name and address of officer to whom the quotation is to be sent are noted above. Any quotation received after the time fixed on the due date is liable to be rejected. The maximum period required for delivery of the articles should also be mentioned. Quotations not stipulating period of firmness and with price variation clause and/or 'subject to prior sale' condition are liable to be rejected.

The acceptance of the quotations will be subject to the following conditions:

- 1. Acceptance of the quotation constitutes a concluded contract. Nevertheless, the successful tenderer must within a fortnight/a month after the acceptance of his quotation furnish 5 per cent of the amount of the contract as security deposit and execute an agreement at his own cost for the satisfactory fulfillment of the contract, if so required.
- Withdrawal from the quotation after it is accepted or failure to supply within a specified time or according to specifications will entail cancellation of the order and purchases being made at the offerers expenses from elsewhere, any loss incurred thereby being payable by the defaulting party. In such an event the Government reserves also the right to remove the defaulter's name from the list of Government suppliers permanently or for a specified number of years.
- Samples, duly listed, should be forwarded if called for under separate cover and the unapproved samples got back as early as possible by the offerers at their own expenses and the Government will in no case be liable for any expenses on account of the value of the samples or their transport charges, etc. In case, the samples are sent by railway; the railway receipt should be sent separately,

Technical Specification

ANNEXURE TO QUOTATION NOTICE No.C1/9487/2023/GMCKLM

Sl.No	Name Digital Newborn weighing machine	
1		
2	Adult weighing machine analouge	2
3	Stadiometer	2
4	Digital Thermometer	10
5	Spacer for Metered dose inhaler	30
6	Baby mask for Metered dose inhaler	100
7	Glucometer	4
8	Sphygmomanometer	. 4
9	Pediatric Stethoscope	2
10	Electronic BP apparatus	1
11	Finger pulse oxymeter 101,500	10

Equipment: Weighing Machine Baby (Electronics)

- 1 Should have 5 gm accuracy.
- 2. Should have minimum measuring capacity 25 gms.
- 3. Should have easy to read LED display.
- 4. Should have measure a maximum weight of 20 kg.
- 5. Should have zero calibration.
- 6. Should hold the measured value irrespective of baby movements.
- 7. Should have electronic damping facility for eliminating the reading fluctuation caused by moving baby
- 8. Weighing pan should be suitable for weighing new born babies and the construction should not allow the baby to slip from the tray.
- 9. The tray should be made of fiber glass/acrylic.
- 10. Should have pan size of at least 50cms length, 20cms width and 8cms height.
- 11. The pan should have facility to measure the length of baby.
- 12. Should operate on input supply 200 to Vac, 50Hz
- 13. Should bear the certification for each machine by department of legal metrology valid in Govt. of Kerala. The tenderer shall submit copy of the model approval issued by Govt. of Kerala, Legal Metrology department along with technical bid.

For PRINCIPAL