

SL NO	ITEM NAME	SPECIFICATION	UNIT	QTY
1	BP Apparatus Cuff	Spare Non-latex Bag Adult 18cm x 50cm, The inflating rubber bag should be capable of withstanding an internal pressure of 300 mmHg without leaking.	Nos	100
2	BP Apparatus Cuff Peadiatric	Spare Non-latex Bag 14cmx21cm, The inflating rubber bag should be capable of withstanding an internal pressure of 300 mmHg without leaking.	Nos	6
3	BP Apparatus Inflator with Air Releasing Valve	Rubber Bulb With Valve, Made out of heavy duty latex, seamless smooth and easy to pump, can use this valve & bulb with most manual blood pressure.	Nos	150
4	BP Apparatus Glass rod Elcometer unit	Made of Borosilicate glass material, Flawless, accurate measurement, clear scale	Nos	5
5	BP Apparatus Glass rod Diamond unit	Made of Borosilicate glass material, Flawless, accurate measurement, clear scale	Nos	5
6	BP Apparatus With Velcro Cover	High quality cotton cover with velcro	Nos	20
7	O2 Cell for Ventilator	Should be an electrogalvanic Oxygen sensor/Cell to monitor Fraction of inspired Oxygen in invasive/non invasive Ventilator. Should monitor the FiO2 from the range between 21% to 100% in the delivered gas by the invasive or non invasive ventilator. Should be compatible to the following models of the ventilator available in GMCH Kollam, 1. Airliquid (extend Xt. Monal T60) 2. Mindray(SV 300) 3. BEL(CV200) Should be factory calibrated and test report of the same shall be submitted. Sample should be provided for item selection to check the compatibility and performance evaluation Should comply with national or international safety standards Warranty- 6 Months	Nos	10
8	Connection Tubes of Ventilator Air Compressor	4mm -16mm Legris pu tube 8*5.5mm , Materials: Polyurethane	meter	5
9	Key Plug for O2, Din Type	BS Type Connector	Nos	10
10	NIBP Cuff (Large) for Multiparameter Monitor	Compatible with Philips, Mindray, Comen, Schiller, BPL, Space Lab Patient Monitors.	Nos	30

*LSR*  
For Principal

Yours faithfully,



(MOTHER TONGUE IS OUR ADMINISTRATIVE LANGUAGE)

**GOVERNMENT MEDICAL COLLEGE**

**PARIPALLY, KOLLAM-691574**

Telephone Office : 0474 2572574, Fax: 0474 2572579 Telephone Principal: 0474-2572572  
Electronic mail: gmckollam@gmail.com Section Email: store.gmckollam@gmail.com

No.C1/9415/2022/GMCKLM

Dated: 05/01/2023

**TENDER NOTICE**

Sealed Competitive Tenders in the Prescribed form are invited for the supply of the following items required for the department of Biomedical Engineering, Government Medical College, Kollam.

Tender should be super scribed with tender number and addressed to the principal, Medical College, Kollam, Pin-691 574. Details can be had from this office on working hours till 5.00 PM

**SCHEDULE**

Sl. No	Tender No	Name of Items	Qty.	Department	Total Cost	D/E	Cost of Tender form (Inclusive vat)		Last date of sale of Tender form	Last date of receipt of tender	Opening date and time
							Original	Duplicate			
1	17/2022	Spares / Accessories for existing Medical Equipmnet (List attached)	Vari ous	Biomedical Engineering	₹1,12,750/-	1%	379	190	19/01/2023 3pm	20/01/2023 2.30pm	20/01/2023 3pm

Tender form will be sent through post if required provided due amount with sufficient postage charge (Rs.50/-) is sent by way of money order in favour of the Principal, Government Medical College, Kollam, Pin-691 574. Money order should reach this office on or before 16/01/2023. No other forms of remittance will be accepted.

FOR PRINCIPAL



“ ഭരണഭാഷ-മാതൃഭാഷ ”

**സർക്കാർ മെഡിക്കൽ കോളേജ്**

പാതിപ്പള്ളി, കൊല്ലം-691574

ഫോൺ: 0474 257 2574, ഫാക്സ്: 0474-2572579, ഇമെയിൽ: പ്രിൻസിപ്പാൾ: 0474-2572572

ഇമെയിൽ: gmcckollam@gmail.com സെക്ഷൻ ഇലക്ട്രോണിക് മെയിൽ: store.gmcckollam@gmail.com

നമ്പർ.സി/9415/2022/ജി.എം.സി. കൊല്ലം

തീയതി. 05/01/2023

**ദർഘാസ് പരസ്യം**

കൊല്ലം സർക്കാർ മെഡിക്കൽ കോളേജിലെ ബയോമെഡിക്കൽ വിഭാഗത്തിലേക്കായി ആവശ്യമായ താഴെപ്പറയുന്ന സാധനങ്ങൾ വിതരണം ചെയ്യുന്നതിന് നിശ്ചിത ഫോറത്തിൽ മുദ്രവച്ച ദർഘാസുകൾ ക്ഷണിക്കുന്നു. ദർഘാസുകൾ ദർഘാസ് നമ്പരോടുകൂടി നിശ്ചിത തീയതിക്കുമുമ്പ് പ്രിൻസിപ്പാൾ, ഗവ.മെഡിക്കൽ കോളേജ്, കൊല്ലം - 691 574 എന്ന വിലാസത്തിൽ എത്തിച്ചിരിക്കേണ്ടതാണ്.

ക്രമ നമ്പർ	ദർഘാസ് നമ്പർ	സാധനങ്ങളുടെ പേര്	എണ്ണം	ഡിപ്പാർട്ടുമെന്റ്	അടങ്കൽ തുക	പിടിയിലായ ശതമാനം	ദർഘാസ് ഫോറത്തിന്റെ വില		ഫോറം സ്വീകരിക്കുന്ന അവസാന തീയതി	ഫോറം തുറക്കുന്ന തീയതി	
							പുഴു	കുഴു			
1	17/2022	സ്പെഷലൈസ് / അക്സസറീസ് ഷോർ എക്സിസ്റ്റിംഗ് മെഡിക്കൽ ഇക്വപ്മെന്റ്സ്	വിവരം	ബയോമെഡിക്കൽ എൻജിനീയറിംഗ്	₹ 1,12,750/-	1%	379	190	19/01/2023 3 മണി	20/01/2023 2.30 മണി	20/01/2023 3 മണി

ഫോറങ്ങൾ തപാലിൽ വേണ്ടവർ വേണ്ടവർ അയയ്ക്കുന്നതിനുവേണ്ട തുക (രൂപ.50/-) അധികമായി അയയ്ക്കേണ്ടതാണ്. ദർഘാസ് ഫോറത്തിന്റെ വില ഓഫീസിൽ നേരിട്ടോ മണിയോർഡർ ആയോ സ്വീകരിക്കുന്നതാണ്. അല്ലാതെയുള്ള എല്ലാ അപേക്ഷകളും നിരസിക്കുന്നതാണ്. ഫോറത്തോടൊപ്പം മുദ്രപത്രങ്ങളോ, മുദ്രപത്രങ്ങളോ അയയ്ക്കാൻ ആവശ്യപ്പെടുകയാണുള്ളതും മേൽവിലാസം വ്യക്തമല്ലാത്തതുമായ മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. മണിയോർഡറുകൾ വഴി ഫോറങ്ങൾ ആവശ്യപ്പെടുകയാണുള്ള അപേക്ഷകൾ ഈ ഓഫീസിൽ 16/01/2023 നു മുമ്പ് കിട്ടിയിരിക്കണം. താമസിച്ച് കിട്ടുന്ന മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. ദർഘാസ് ഫോറം കൈമാറാൻ പാടില്ലാത്തതും ഒരിക്കൽ അടച്ച തുക തിരിച്ചുകിട്ടുന്നതുമല്ല. താമസിച്ച് കിട്ടുന്ന ദർഘാസുകൾ യാതൊരു കാരണവശാലും സ്വീകരിക്കുന്നതല്ല. കൂടുതൽ വിവരങ്ങൾ പ്രവൃത്തി ദിവസങ്ങളിൽ ഈ ആഫീസിൽ നിന്നും ലഭിക്കുന്നതാണ്.

*(Signature)*  
പ്രിൻസിപ്പാളിനുവേണ്ടി